



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JULY 2023



Constituent Group

Chemistry III

Lab No

Date of Result Entered :

10635

PT item

Lyophilized human serum based

Date of Report Published :

07/08/2023

18/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi Automatic)	1876	109.35	5.55	6.07	106.4 mg/dL	-0.49	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi Automatic / Fully Automatic)	1850	5.83	6.00	0.35	5.84 mg/dL	0.03	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi Automatic / Fully Automatic)	1776	113.00	4.12	4.66	115.4 mg/dL	0.52	0.22

SDI Range		Interpretation
Within -1.00 to +1.00		Excellent.
Within ± 1.01 to ± 2.00		Good.

home.cmcvellore.ac.in/clinqc/ViewMonthlySummary/printResult_new.aspx?resDelid=267&noSamples=1&month=7&year=2023&groupid=9&yearGroup=110

THE MEDICAL OFFICER
GOVT. PRIMARY HEALTH CENTRE
KATCHERITHAL AVANIPURAM
THOOTHUKUDI DIST 1/2