



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME



Constituent Group: Chemistry III
PT item: Lyophilized human serum based

Date of Result Entered: 04/07/2023
Date of Report Published:

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	307.2 mg/dL	1.39	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.45 mg/dL	-0.70	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	116.9 mg/dL	0.63	0.24

SDI Range

Excellent.

Good.

Within -1.00 to +1.00

Within ±1.01 to ±2.00

Interpretation

MEDICAL OFFICER
GOVT. PRIMARY HEALTH CE
KATCRERITHALVAIPUR
THOOTHUKUDI DIST

cvelllore.ac.in/clinqc/ViewMonthlySummary/printResult_new.aspx?resDetId=266&noSamples=1&month=6&year=2023&groupId=9&yearGroup=110