



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME



Lab Name: Govt. PNC, Katchadavalapuram
Constituent Group: Chemistry III
PT Item: Lyophilized human serum based

Lab No: 19105/2023
Date of Result Entered: 05/06/2023
Date of Report Published:

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	176 mg/dL	-0.52	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi Automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.82 mg/dL	0.45	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi Automatic / Fully Automatic)	1798	114.60	5.08	5.83	126.6 mg/dL	2.06	0.27

SDI Range		Interpretation
Within -1.00 to +1.00		Excellent.
Within ± 1.01 to ± 2.00		Good.

MEDICAL OFFICER
GOVT. PRIMARY HEALTH CENTRE
KATTCHERITHALANAPURAM
THOOTHUKUDI DIST