



CHRISTIAN MEDICAL COLLEGE

MONTHLY SUMMARY REPORT - APRIL 2023



Lab Name: Govt. PHC, Katchaidalavaipuram

Lab No: 10635

Constituent Group: Chemistry III

Date of Result Entered: 20/04/2023

PT item: Lyophilized human serum based

Date of Report Published: 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Year Val.	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	118.9 mg/dL	-0.08	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.34 mg/dL	1.03	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	110.5 mg/dL	-0.19	0.28

SDI Range		Interpretation
Within -1.00 to +1.00		Excellent.
Within ±1.01 to ±2.00		Good.