

CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/12873		MEDICAL DEVICES		Page No:1 of 1
Issue Date	06-10-2023			
SRF No & Date	SRF/23/00638-0001 & 05-10-2023			
Receipt Date	05-10-2023			
Calibration Date	05-10-2023			
Calibration Due	04-10-2024			
Customer Name & Address				
GOVERNMENT PRIMARY HEALTH CENTRE, E.PUDUPATTI-625604, THENI DISTRICT.				
Details of Device Under Calibration (DUC)				
Description	: SEMI AUTO ANALYZER	Make & Model	: ROBONIK & PRIETEST TOUCH	
Range	: MULTI	Sr. No	: AT1640816RBK	
Resolution	: MULTI	Identification No	: NA	
DUC Condition	: SATISFACTORY	Location	: LABORATORY	
Environmental Conditions & Calibration Procedure Details				
Environmental Details	Temperature:25.6 ° C	Relative Humidity	52% RH	
Calibration Procedure No	SBS/CP/MD/20	Calibration done at	ONSITE	
Reference Standards Details				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

ELECTRICAL SAFETY

RESULTS

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resisitance	Measured values in MΩ	Uncertainty in % (±)
	>20MΩ	79	13.92
2	Earth Leakage	Measured values in µA	Uncertainty in % (±)
	<5000µAfor B,BF,CF	209	7.3
3	Enclosure Leakage	Measured values in µA	Uncertainty in % (±)
	<500µAfor B,BF,CF	229	7.2

REMARKS

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

Authorised Signatory


(Calibration Engineer)
P.PRASANNA


Technical Manager
C.SHANMUGARAJ

Chief Executive



SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.



CALIBRATION CERTIFICATE

CERTIFICATE NO:	SBS/CL/23/12874	MECHANICAL	Page No:1 of 1
Issue Date	06-10-2023		
SRF No & Date	SRF/23/00638-0002 & 05-10-2023		
Receipt Date	05-10-2023		
Calibration Date	05-10-2023		
Calibration Due	04-10-2024		

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE,
E.PUDUPATTI-625604, THENI DISTRICT.

Details of Device Under Calibration (DUC)

Description	: CENTRIFUGE	Make & Model	: LABOTECH & BDI-152
Range	: 3500 RPM	Sr. No	: NA
Resolution	: 1 RPM	Identification No	: NA
DUC Condition	: Satisfactory	Location	: LABORATORY

Environmental Conditions & Standard Operating Procedure Details

Environmental Details	Temperature: 24.3 °C	Relative Humidity	54 % Rh
Calibration Procedure No	SBS/CP/ML/04	Calibration done at	ONSITE

Reference Standards Details

S.No	Description	Make/ SI No:	Certificate No	Validity
1	Digital Tachometer	Lutron & DT-2238	JRPM-CCTR-A&S-2022-0036	03-10-2023

CALIBRATION RESULTS

S.No	DEVICE UNDER CALIBRATION READINGS	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY (±)
	RPM	RPM	RPM	%
1	1000	999.8	0.2	4.2
2	2000	1998	2	4.2
3	3000	2998	2	4.2

REMARKS

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By


(Calibration Engineer)
P.PRASANNA



Authorised Signatory


Quality Manager
(C.SIVABALAN)

Chief Executive

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.