



VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Belarahalli

[Details About Robust Analysis](#)

Lab No 12120

[Detail About Monthly Summary](#)

Month April

[Detail about SDI](#)

Year 2023

Constituent Group Chemistry III

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 20/04/2023

Date of Report Published : 06/05/2023

| Sl.No | Analyte                        | Method / Principle Name               | Analyzer Name                                | No of Participants | DV     | Participants |      | Your Value | SDI  | U    |
|-------|--------------------------------|---------------------------------------|----------------------------------------------|--------------------|--------|--------------|------|------------|------|------|
|       |                                |                                       |                                              |                    |        | CV           | SD   |            |      |      |
| 1     | <a href="#">GLUCOSE III</a>    | GOD-POD III                           | Robonik ( Semi automatic / Fully Automatic ) | 1832               | 119.45 | 6.04         | 7.21 | 122 mg/dL  | 0.35 | 0.34 |
| 2     | <a href="#">CREATININE III</a> | Jaffes Kinetic - Alkaline Picrate III | Robonik ( Semi automatic / Fully Automatic ) | 1820               | 2.16   | 8.09         | 0.18 | 2.2 mg/dL  | 0.23 | 0.01 |
| 3     | <a href="#">CHOLESTROL III</a> | CHOD-PAP III                          | Robonik ( Semi automatic / Fully Automatic ) | 1750               | 111.61 | 5.19         | 5.79 | 112 mg/dL  | 0.07 | 0.28 |

| SDI Range             | Interpretation                           |
|-----------------------|------------------------------------------|
| Within -1.00 to +1.00 | Excellent.                               |
| Within ±1.01 to ±2.00 | Good.                                    |
| Within ±2.01 to ±2.99 | Accept with caution. Warning Signal.     |
| Beyond ±3.0           | Unacceptable performance. Action Signal. |

|                                |                                 |                                    |                                          |
|--------------------------------|---------------------------------|------------------------------------|------------------------------------------|
| Links                          |                                 |                                    |                                          |
| <a href="#">Modify Address</a> | <a href="#">Modify Methods</a>  | <a href="#">Submit Lab Results</a> | <a href="#">View Lab Monthly Summary</a> |
| <a href="#">Yearly Summary</a> | <a href="#">Change Password</a> | <a href="#">Home</a>               | <a href="#">Sign-out</a>                 |