

[VIEW LAB MONTHLY SUMMARY](#)

Lab Name            Govt. PHC, Belarahalli

[Details About Robust Analysis](#)

Lab No                12120

[Detail About Monthly Summary](#)

Month                March

[Detail about SDI](#)

Year                  2023

Constituent Group    Chemistry III

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 20/03/2023

Date of Report Published : 04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	<a href="#">GLUCOSE III</a>	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1878	91.48	7.46	6.82	93 mg/dL	0.22	0.31
2	<a href="#">CREATININE III</a>	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1851	1.33	9.08	0.12	1.4 mg/dL	0.58	0.01
3	<a href="#">CHOLESTROL III</a>	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1761	110.93	5.23	5.80	105 mg/dL	-1.02	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>