

# HOPE ULTRA SYSTEMS

No: 2/79, MAIN ROAD, MAMSAPURAM (PO),  
SIVAKASI, VIRUDHUNAGAR-626124.  
E-mail : [hopeultrasystems@gmail.com](mailto:hopeultrasystems@gmail.com)

GSTIN. 33DLMPK3057D1Z9  
Mobile: +919487402761, +919715223192.

## CALIBRATION CERTIFICATE

Requested By  
**GPHC, M. SUBBULAPURAM,.**

<b>Scope:</b>		Asset no	: PHC/MSP-LAB/003
Equipment Name	: Micropipette	Calibration Done at	: customer place
Make	: BIOHIT	Calibration Date	: 09/09/2023
Model	: 10-100 $\mu$ l	Next call due on	: 08/09/2024
Sl no	: 11504418		

### TEST RESULTS Volume $\mu$ l

Sl. No.	DUC reading( $\mu$ l)	Standard Reading ( $\mu$ l)		Correction ( $\mu$ l)
		Trail 1	Trail 2	
01	10	10.0	10.0	+0
02	20	20.0	20.0	+0
03	50	50.0	50.0	+0
04	80	80.0	80.0	+0
05	100	100.0	100.0	+0

Error Acceptance : Not Applicable (user defined)

Equipment Status : Working in Good condition


Materials used for testing and calibration:

- Electronic weighing scale(EWC/SF400C)

Note:

- Calibration points were selected as per customer request.
- Each reading was taken by repeatability.
- Results reported are valid at the time of and under the stated conditions of measurements.

*Hope Ultra Systems*

  
Authorized signatory