

HOPE ULTRA SYSTEMS

No: 2/79, MAIN ROAD, MAMSAPURAM (PO),
SIVAKASI, VIRUDHUNAGAR-626124.

E-mail : hopeultrasystems@gmail.com

GSTIN. 33DLMPK3057D1Z9

Mobile: +919487402761, +919715223192.

CALIBRATION CERTIFICATE

Requested By

GPHC, M. SUBBULAPURAM.

Scope:

Equipment Name : CENTRIFUGE	Asset no : PHC/MSP-LAB/006
Make : LABOTECH	Calibration Done at : customer place
Model : BD-153	Calibration Date : 09/09/2023
Sl no : 76694	Next call due on : 08/09/2024

Sl. No.	DUC reading(RPM)	Standard Reading (RPM)	Correction (RPM)
01	Position 1	2010	+0
02	Position 2	2150	+0
03	Position 3	2670	+0
04	Position 4	3330	+0
05	Position 5	3850	+0

Error Acceptance : Not Applicable (user defined)

Equipment Status : Working in Good condition


Materials used for testing and calibration:

- Digital Multimeter-Mastech-MS830L
- Digital Taco meter-DT-2234C/Max tech

Note:

- Calibration points were selected as per customer request.
- Each reading was taken by repeatability.
- Results reported are valid at the time of and under the stated conditions of measurements.

Hope Ultra Systems


Authorized signatory