



CHRISTIAN MEDICAL COLLEGE

The Pioneer of the National Class - 1969

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. UPHC, Appavu Nagar
 Lab No: 10923
 Month: July
 Year: 2023
 Constituent Group: Chemistry III

[Details About Robust Analysis](#)
[Detail About Monthly Summary](#)
[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 19/07/2023

Date of Report Published : 07/08/2023

Sl. No.	Analyte	Method / Reagent	Instrument	No. of Tests	Mean	SD	CV	Target	SDI	CVI
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	112 mg/dL	0.44	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.63	6.00	0.35	5.9 mg/dL	0.20	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	121 mg/dL	1.72	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links
View Graphical Data
View Analyser Result
View Report
View Summary

Site Maintained by **Swiftware Solutions**

Best Viewed at pixels 1024 X 768

[Signature]
 MEDICAL OFFICER
 URBAN PRIMARY HEALTH CENTER
 APPAVU NAGAR, HOSUR,
 Krishnagiri District.