



CHRISTIAN MEDICAL COLLEGE
 DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - SEPTEMBER 2023

Lab Name: Govt. PHC, Pudukattai,W
 Constituent Group: Chemistry III
 PT Item: Lyophilized human serum based

Lab No: 10673
 Date of Result Entered: 20/09/2023
 Date of Report Published: 05/10/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1870	310.86	4.82	14.98	317 mg/dL	0.41	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1882	1.29	8.14	0.10	1.2 mg/dL	-0.86	0.00
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1783	113.62	3.80	4.32	114 mg/dL	0.09	0.20
SDI Range										
Within -1.00 to +1.00					Excellent.					
Within ±1.01 to ±2.00					Good.					
Within ±2.01 to ±2.99					Accept with caution. Warning Signal.					
Beyond ±3.0					Unacceptable performance. Action Signal.					

LAB ADDRESS :
 Govt. PHC, Pudukattai,W
 CLINICAL LABORATORY , VIRUDHUNAGAR District ,
 Pudukattai,W
 TAMILNADU626130

Coordinator Contact Details:
 Email:clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christodoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****