

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. UPHC, Appavu Nagar
 Lab No: 10923
 Month: April
 Year: 2023
 Constituent Group: Chemistry III

[Details About Robust Analysis](#)
[Detail About Monthly Summary](#)
[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)


Date of Result Entered : 18/04/2023

Date of Report Published : 06/05/2023

Sl. No.	Parameter	Method	Result	Unit	Reference Range	SDI	CV	SI	SI
1	GLUCOSE III	GOD-POD III	1832	mg/dL	119.45 - 6.04 - 7.21	132	1.74	0.34	
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	1820	mg/dL	2.16 - 8.09 - 0.18	2.4	1.37	0.01	
3	CHOLESTROL III	CHOD-PAP III	1750	mg/dL	111.61 - 5.19 - 5.79	112	0.07	0.28	

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within +1.01 to +2.00	Good.
Within +2.01 to +2.99	Accept with caution. Warning Signal
Beyond +3.0	Unacceptable performance. Action Signal

Links	Parameter	Result	Unit
GLUCOSE III	1832	mg/dL	
CREATININE III	1820	mg/dL	
CHOLESTROL III	1750	mg/dL	

Site Maintained by
 **Swiftware Solutions**

Best Viewed at pixels 1024 x 768


MEDICAL OFFICER
 URBAN PRIMARY HEALTH CENTER
 APPAVU NAGAR, HOSUR,
 Krishnagiri District.