

EW LAB MONTHLY SUMMARY

Lab Name: Govt. UPHC, Appavu Nagar
 Lab No: 10923
 Month: August
 Year: 2023
 Constituent Group: Chemistry III

Details About Robust Analysis
 Detail About Monthly Summary
 Detail about SDI

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)


Date of Result Entered : 16/08/2023

Date of Report Published : 01/09/2023

Sl. No.	Analyte	Method / Principle	Instrument	Q10	Q20	Q30	Q40	Q50	Q60	Q70	Q80	Q90
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	189 mg/dL	1.49	0.42		
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.8 mg/dL	0.34	0.01		
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	126 mg/dL	2.63	0.22		

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Site Maintained by
 **Swiftware Solutions**

Best Viewed at pixels 1024 x 768


MEDICAL OFFICER
 URBAN PRIMARY HEALTH CENTER
 APPAVU NAGAR, HOSUR,
 Krishnagiri District.