

			CAL	<b>BRAT</b>	10	N CER	TIFIC	A	TE_	
CERTIFICATE NO: SBS/CL/23/14883				MECHANICAL				L	Page No:1 of 1	
Issue Date					17-10-2023					
SRF No & Date					SRF/23/00904-0002 & 16-10-2023					
Receipt Date					18-10-2023					
Calibration Date					16-10-2023					
Calibration Due					15-10-2024					
Custon	ner Name &	Addres	3						2	
GOVE	RNMENT P	RIMAR	Y HEALTH CENTRE,							
HOSU	R-635121,F	RISHN	AGIRI DISTRICT.							
			De	talls of De	vice i	Under Callbi	ration (D	JC)		
Descrip	otion	:	CENTRIFUGE		Make	& Model	:		NA & NA	
Range : 35			3500 RP	M	Sr. No				5240	
Resolution :			1 RP	M	Identification No				NA	
DUC C	ondition	<u>:</u>	Satisfactory		Loca				LABORATORY	
			Environmental	Condition	8 & S	tandard Ope	rating P	roce	dure Details	
Environmental Details Temperatur				ture: 25.5%	25.5°C Relative Hu		midity 52%Rh		52%Rh	
Calibration Procedure No SBS/CP/ML				VIL/04	/04 Calibration		done at ONSITE		ONSITE	
				Refer	ence (	Standards D	etails			
S.No	No Description			Make/	Make/ SI No:		Certificate No			Validit
1	Digital Tachometer			LINE S	LINE SEIKI / 175-0034V		JRPM-CCTR-A&S-2023-0013			09-06-20

## **CALIBRATION RESULTS**

S.No	DEVICE UNDER CALIBRATION	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY	
	RPM	RPM	RPM	%	
1	1000	998.9	1.1	4.2	
2	2000	1999.5	0.5	4.2	
3	3000	2999.5	0.5	4.2	

## REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

(Calibration Engineer)

K.SATHYAMOORTHY



Authorised Sig	natory
(C.SIVABALAN)	(Chief Executive