

				CAL	IBRAT	101	I CERT	IFIC	A7	<u>E</u>		
CER	CERTIFICATE NO: SBS/CL/23/13405				MEDICAL DEVICES				CES	Page	No:1 of 1	
Issue	Date					07-10-2023						
SRF	No & Date					SRF/23/00741-0001 & 07-10-2023						
Rece	ipt Date					07-10-2023						
Calib	ration Date					07-10	0-2023					
Calib	ration Due					06-10	0-2024					
Cust	omer Name	& Add	ress									
GOV	ERNMENT	PRIMA	RY HEA	LTH CENTRE	Ξ,							
OLAF	PALAYAM-6	38052,	KARUR	DISTRICT.								
					Details of De	vice U	Inder Calibrat	ion (DU	-			
Description : SE		SEMI A	AUTO ANALYZER		Make & Model		:		ROBONIK & PRIETEST TOUCH		4	
Range :		MULTI	ILTI		Sr. No		:		AT2190211RBK			
Resolution : MULT		MULTI			Identi	fication No	:		NA			
DUC Condition : SATISFACTORY		Location		:		LABORATORY						
				Environr	nental Condi	tions	& Calibration	Procedu	ire De	tails		
Environmental Details Temperature:25.6 ° C			25.6 ° C	Relative Humidity			52% RH					
Calibration Procedure No			SBS/CP/MD/20		Calibration do		one at		ONSITE			
	110,0				Refere	nce S	tandards Deta	ails				
S.No Description Make/ SI N			Make/ SI No:	o:		Certificate No				Validity		
1 Electrical Safety Analyser Rig			Rigel Medical & 44L-1059		M-230809-16-4				10-08-2024			

ELECTRICAL SAFETY

_			
P	FSI	11	TS

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resisitance	Measured values in $M\Omega$	Uncertainty in % (±)
	>20MΩ	92	13.92
2	Earth Leakage	Measured values in µA	Uncertainty in % (±)
	<5000µAfor B,BF,CF	172	7.5
3	Enclosure Leakage	Measured values in µA	Uncertainty in % (±)
	<500µAfor B,BF,CF	221	7.2

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

 Calibrated By

 Authorised Signatory

KSty

7

(Calibration Engineer) K.SATHYAMOORTHY



Technical Manager
C.SHANMUGARAJ

Chief Executive	9
-----------------	---

SUNSHINE BIOMEDICAL SOLUTIONS