

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Arumuganeri	Details About Robust Analysis								
Lab No	12985	Detail About Monthly Summary								
Month	March	Detail about SDI								
Year	2023									
Constituent Group	Chemistry III									
Click on the analyte to view Graphical Data						All Analyser Result		Print		
Date of Result Entered : 19/03/2023										
Date of Report Published : 04/04/2023										
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	80 mg/dL	-1.68	0.31
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	175	55.20	11.70	6.46	56 mg/dL	0.12	0.98
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.1 mg/dL	-1.90	0.01
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	110 mg/dL	-0.16	0.28
SDI Range			Interpretation							
Within -1.00 to +1.00			Excellent.							
Within ±1.01 to ±2.00			Good.							
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.							
Beyond ±3.0			Unacceptable performance. Action Signal.							

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out