

# Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No. 770/1, Iothi Complex 2nd Floor, Alagar Kovil Main Road, Surveyor Colony,  
Madurai - 625 007, Tamilnadu, India Mobile : +91 94434 74610



GSTIN: 33AD5PA9153Q1ZA • DL No. TN-12-208-00398, TN-12-218-00398

## CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Government Primary Health Centre, Salai Puthur,	Certificate No:	AMS/HOSP/289/3298
		Issuing Date:	28/03/2023
		Page No:	1 of 1

### Details of Device under Calibration (DUC):

DUC:	Micropipette	Id. No:	GPHC/LAB/VMP/001
Make:	Finnpipette F73	Department:	Laboratory
Model:	Thermo	Calibrated On:	27/03/2023
Sl. No:	RW10524	Calibration Due Suggested:	26/03/2024
Type:	Variable	DUC Condition on Receipt:	OK
Range & Least Count:	10 – 100µl		

### Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electronic Weighing Scale	EWC/SF400C	PC/01/23/ATL/0512-02	31/01/2024
2	Thermo Hygrometer	Beetech/088	PC/01/23/ATL/0210-02	31/01/2024

Room Temperature: 25°C±2 Room Humidity: 30-60%

(At the time of Calibration)

### Test Results Volume µl:

SL. No.	DUC Reading (µl)	Standard Reading (µl)		Correction (µl)
		Trial 1	Trial 2	
1	10.0	10.00	10.00	00.00
2	20.0	20.00	20.00	00.00
2	50.0	50.00	50.00	+00.00
3	100.0	105.00	105.00	+05.00

Allowable Tolerance: User Defined

### Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

### Calibrated By:

(Calibration Engineer)

aceagm@gmail.com



ams@healthcareindia@gmail.com



### Approved by:

(Technical Manager)

94434 74610, 93455 56148

# Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No. 770/3, Jothi Complex 2nd Floor, Alagar Kovil Main Road, Surveyor Colony,  
Madurai - 625 007, Tamilnadu, India Mobile : +91 94434 74610



GSTIN: 33AD5PA9153Q1ZA • DE No. TN-12-20B-00398, TN-12-21B-00398

## CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Government Primary Health Centre, Salai Puthur,	Certificate No:	AMS/HOSP/289/3299
		Issuing Date:	28/03/2023
		Page No:	1 of 1

### Details of Device under Calibration (DUC):

DUC:	Micropipette	Id. No:	GPHC/LAB/VMP/002
Make:	Finnpipette F <sup>3</sup>	Department:	Laboratory
Model:	Thermo	Calibrated On:	27/03/2023
Sl. No:	RW13668	Calibration Due Suggested:	26/03/2024
Type:	Variable	DUC Condition on Receipt:	OK
Range & Least Count:	100 – 1000µl		

### Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electronic Weighing Scale	EWC/SF400C	PC/01/23/ATL/0512-02	31/01/2024
2	Thermo Hygrometer	Beetech/088	PC/01/23/ATL/0210-02	31/01/2024

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

### Test Results Volume µl:

SL. No.	DUC Reading (µl)	Standard Reading (µl)		Correction (µl)
		Trial 1	Trial 2	
1	100.0	110.00	110.00	+10.00
2	200.0	210.00	210.00	+10.00
3	500.0	500.00	500.00	+00.00
4	800.0	800.00	800.00	+00.00
5	1000.0	990.00	990.00	-10.00

Allowable Tolerance: User Defined

### Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test(i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

### Calibrated By:

(Calibration Engineer)



### Approved by:

(Technical Manager)



# Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No. 770/1, Jothi Complex 2nd Floor, Alagar Kovil Main Road, Surveyor Colony,  
Madurai - 625 007, Tamilnadu, India Mobile : +91 94434 74610



GSTIN: 33ADSPA9153Q1ZA • DL No: TN-12-208-00398, TN-12-218-00398

## CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Government Primary Health Centre, Salai Puthur,	Certificate No:	AMS/HOSP/289/3300
		Issuing Date:	28/03/2023
		Page No:	1 of 1

### Details of Device under Calibration (DUC):

DUC:	Micropipette	Id. No:	GPHC/LAB/VMP/003
Make:	Microlux	Department:	Laboratory
Model:	---	Calibrated On:	27/03/2023
Sl. No:	---	Calibration Due Suggested:	26/03/2024
Type:	Variable	DUC Condition on Receipt:	OK
Range & Least Count:	10 - 1000µl		

### Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electronic Weighing Scale	EWC/SF400C	PC/01/23/ATL/0512-02	31/01/2024
2	Thermo Hygrometer	Beetech/088	PC/01/23/ATL/0210-02	31/01/2024

Room Temperature: 25°C±2      Room Humidity: 30-60%      (At the time of Calibration)

### Test Results Volume µl:

SL. No.	DUC Reading (µl)	Standard Reading (µl)		Correction (µl)
		Trial 1	Trial 2	
1	10.0	10.00	10.00	00.00
2	20.0	20.00	20.00	00.00
2	50.0	50.00	50.00	+00.00
3	100.0	105.00	105.00	+05.00
4	200.0	201.00	201.00	+01.00
5	500.0	503.00	503.00	+03.00
6	1000.0	1005.00	1005.00	+05.00

Allowable Tolerance: User Defined



### Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test(i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:  
(Calibration Engineer)



Approved by:  
(Technical Manager)