

**T. ARUN PRASAD, M.Tech. MBA. (Ph.D)**  
Managing Director  
Consultant Biomedical Engineer

GSTIN: 33APLPA 1828A1ZM

### Calibration Protocol & Report for Medical Equipment

Equipment: **CENTRIFUGE**  
Make:REMI  
Model:REMI

Protect Class/Type: Risk class 1  
S.No :GPHC/RPM-01  
Location : **MEKALACHINNAMPALLI-UPHC**

S.no	Measurement Parameters	Status	Remarks
I	<b>Electrical test</b> (a). Safety Standards (range: 210-240 V) (b). Equipotential ground &Earthing (c).Risk of shock/fire Hazards, if any (d). Chassis&Insulation Resistance (e). Earth & Chassis Leakage Current (f). Battery condition & charging status	230 volts 0 to 5 volts no zero ohms NA	Stabliser/ups  Wih in ranges
II	<b>Physical Condition Test</b> (a). Cabinet & Visible damages, if any (b). Spo2 probe, clip,extn cable, Power cord & accessories (c). Risk of mechanical damages, if any)	No ok No	Checked ok
III	<b>Performance Test</b> (a). On/off switch (b).Speed regulator switch & functional tests (c). Running test (d).Abnormalities (e). Display checkup (f). Keypad/indicators & connectors	Normal Ok accurate ok Ok Well	Perfect Perfect Ok
IV	<b>Calibration Test</b> a). Vibration test (b). Speed regulation gradual hike/ (c). Running test with load (d). Display checkup (f). timer test	Done ok	Perfect
<b>Formulated By:</b> Einstein Research Laboratory		<b>Issued By:</b>	<b>User Department:</b>
<b>Endorsed By:</b> Research team head			
Reviewed On : <b>20.10.2023</b> Next Review: <b>19.04.2024</b>			