



Care



CALIBRATION REPORT

Report No : HBMS/CAL/23-24		Calibration Date: 26.09.2023
		Calibration Valid: 25.03.2024
I.1.CUSTOMER DETAILS:		
Name and Address of the organization		M/S. CORPORATION HOSPITAL, UPHC, SINGANALLUR , COIMBATORE.
I.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):		
EQUIPMENT NAME		MICRO PIPETTE
i	Manufactured by	THERMO SCIENTIFIC
ii	Model	FINEPIPTTE – F3
iii	Value	100 µl -1000 µl
iv	Serial No	PW02569

Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Calibration test Report.

Number of measurement	Channel 1 (Volume)		
	10µl	50µl	100µl
1	100.02	500.06	1000.09
2	100.04	500.04	1000.08
3	100.05	500.02	1000.06
4	100.02	500.02	1000.08
5	100.01	500.02	1000.09
6	100.03	500.07	1000.7

For Heart Beat Medical Systems,

(Signature)
(Authorized signatory)



No: TN/CBE/MD42/00016
N: 33AAJFH1833M1ZB

99400 28355, 9751206106
88833 15662

No.42/1, Navaindia Road, Peelamedu,
CBE – 04, hbmedicalsystem@gmail.com