

System QC protocol

EC 90Analyzer

Serial number of product	4289-2-202208	Manufacturing Date	09/2022
SW version	1.11.3	Test performed by	Burianová
EC cartridge lot	2111139	Test performed at	2.9.2022

No.	Performance content	Limit/Criterion	Sample	Notice	Pass/Fail
1.	Final QC instrument	valid protocol	-	-	✓
Basic QC measurements					
2.	QC measurement performance	less than 8,3% of all values are out of target range	Erba Norm	100 %	✓
3.			Lot:	2105150B	
			Erba Path	100 %	✓
			Lot:	2203101	
Conclusive system check					
4.	SW check	all tasks could be done per SW, no SW-freeze	-	-	✓
5.	Mechanical check	no mechanical damage detectable	-	-	✓
6.	Put into operation	no additional maintenance for put into operation required	-	-	✓
7.	Closing activities	Cleaning, visual tube check, shutdown	-	-	✓

Disposition	Name	Date	Signature
Pass/Fail	Burianová	2.9.2022	