



S.M. ENTERPRISES

27th
Years of Service

(Calibration and Testing of Electrical, Electronic & Precision Instruments)

Office/Cal. Lab : Plot No. 12, "Ganesh Niwas", Sanjeevaiah Nagar, Co-operative Society, Besides Tadbund Hanuman Temple, Sikh Village, Secunderabad - 500 009, T.S. INDIA ☎: (040) 27840697, 48541008
Cell: 9848020798, 9948291008, 9948292008, 9948293008, 9948294008 Email: msudheer_98@yahoo.com, msudheer_98@smentp.org, SME_003@YAHOO.IN, url : www.smentp.org

CALIBRATION CERTIFICATE

CALIBRATION ON 09/10/2023	CALIBRATION DUE 08/10/2024
1. Calibration Report Ref. No. & Date :	SME/SUJ/DSW/CAL-010/2023;Dt:09/10/2023
2. Requested by : (Name and address of the organization)	M/s. PPTCTC GH RIMS Srikakulam GH RIMS Srikakulam
3. Instrument Name	Digital Timer
4. Instrument Make	Acetecq / JS -118
5. Instrument ID. No:	TIM/ICTC/010
6. Range :	0 to 99 min
7. Accuracy :	+/- 1 % FSD
8. Environmental Condition	Temp/RH : 28 +/- 2 Deg C; R.H: 65 +/- 5%
9. Test parameter & condition :	Time V/S Time (Comparison Method)
10. Standard Instruments used for Calibration:	Time Interval Meter
11. Traceability Certificate Ref. No.	C-230613-5-1, Dt:22/06/2023, DueDt:22/06/2024

TEST RESULTS


SI No	Standard Reading Of Time in Master Digital Stop Watch	Observed Reading Of Time in Digital Stop Watch	Error Value (+/-)
	Hr:min:sec:ms	min:sec	Hr:min:sec:ms
1	00:01:00:23	01:00	00:00:00:23
2	00:10:00:33	10:00	00:00:00:33
3	00:30:00:32	30:00	00:00:00:32
4	01:50:00:22	90:00	00:00:00:22

PLEASE NOTE: THE INSTRUMENTS ERROR FOUND TO BE WITHIN THE SPECIFIED ACCURACY LIMIT AND THIS CERTIFICATE VALID FOR PERIOD OF ONE YEAR FROM THE DATE OF ISSUE.

CALIBRATED BY
S. M. Enterprises

ENGINEER INCHARGE

VERIFIED BY
M/s. PPTCTC GH RIMS Srikakulam



GSTIN : 36AFVPS4924F1ZW Service Tax No. : AFVPS4924FST001 PAN No. : AFVPS4924F
NABL APPROVED LAB (CC-2471) / Approved Laboratory of ISO 9001-2015
TEST EQUIPT. / STANDARDS USED ARE TRACEABLE TO NATIONAL STANDARDS