

**CALIBRATION REPORT**

Report No: HBMS/CAL/23-24		Calibration Date: 26.10.2023
		Calibration Valid: 25.04.2024
<b>I.1. CUSTOMER DETAILS:</b>		
Name and Address of the organization		M/s, PRIMARY HEALTH CENTRE, ANAIMALAI.
<b>I.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>		
EQUIPMENT NAME		PIPETTE
I	Manufactured by	ERBA
II	Model	-----
III	Value	1000µl
IV	Serial No	PA535866


**Calibration standards used :**


Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2023	06-OCT-2024

Details of the machine: machine calibrated and checked output.

Number of measurement	Channel 1 (Volume)
	1000µl
1	1000.05
2	1000.06
3	1000.01
4	999.97
5	1000.01
6	1000.02

For Heart Beat Medical Systems,

  
(Authorized signatory)





CALIBRATION REPORT	
Report No: HBMS/CAL/23-24	Calibration Date: 26.10.2023
	Calibration Valid: 25.04.2024
<b>1.1. CUSTOMER DETAILS:</b>	
Name and Address of the organization	M/s, PRIMARY HEALTH CENTRE, ANAIMALAI.
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>	
EQUIPMENT NAME	PIPETTE
I Manufactured by	ERBA
II Model	---
III Value	100µl
IV Serial No	PA 534612

**Calibration standards used :**

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2023	06-OCT-2024

**Details of the machine:** machine calibrated and checked output.

Number of measurement	Channel I (Volume)
	100µl
1	100.05
2	100.06
3	100.01
4	99.97
5	100.01
6	100.02

For Heart Beat Medical Systems,

*S. Anurag*  
(Authorized signatory)

