

# MK BEST CALIBRATION SERVICES



No. 27, F-2, 1st Floor, 2nd Street, Varalakshmi Nagar, Maduravoyal,  
(Opp. MGR Engineering College), Chennai - 600 095.  
Ph.: 044 - 23780211, Cell : 93802 66480 / 86958 18108 / 90032 77250  
E-mail: mkbestcalibration@gmail.com, www.mkbestcalibrationservices.com



## CERTIFICATE OF CALIBRATION



FF/7.8/01

<b>Certificate No</b>	: MKBN/23/09/0416-001	<b>Date of Issue</b>	: 25.09.2023
<b>Date of Calibration</b>	: 22.09.2023	<b>Due Date of Calibration</b>	: 21.09.2024

CUSTOMER INFORMATION		UNIT UNDER CALIBRATION	
M/S. KB DIAGNOSTICS, VANDAVASI.		Description	CENTRIFUGE
		Range	3000 rpm
		Resolution	10 rpm
		Make	JIANGSU XINKANG MEDICAL INSTRUMENT CO., LTD
		Model	ELECTRA-BL-12B
		Identification No	KBD/CB/GEN/01
		Serial No	202207225-12
		Location	LAB

S.No	Description	Id.No/Sl. No	Traceability	Certificate No	Validity
01	Digital Tachometer	MK/CAL-222	National Standards	CRMTL/01/422103978-A4	02.12.2023

ENVIRONMENTAL CONDITIONS		REFERENCE STANDARD	
Temperature	23.2°C	Uncertainty ± °C	1.57
Humidity	59 % RH		

### CALIBRATION RESULTS

S.NO	Standard Reading rpm	UUC Readings rpm	Error rpm
1	507	500	-7
2	1015	1000	-15
3	2028	2000	-28
4	3034	3000	-34

**Remarks :**

- The above UUC was Calibrated at Site.
- The Calibration Certificate Shall not be Reproduced Except In Full, Without Written Approval Of The Laboratory.
- The Recalibration Interval is Determined based on the User requirement.
- The Results Stated In This Certificate Relate Only to the Item Calibrated.
- The above Instrument used for Medical Purpose.

Calibrated by

*S. Chandrabose*  
S.Chandrabose  
(Calibration Engineer)

x-x-x-x End Of Certificate x-x-x-x



Authorised by

*L. Magesh*  
L.Magesh  
(MD/QM)