

# MK BEST CALIBRATION SERVICES



NABL ACCREDITED CALIBRATION LABORATORY AS PER ISO/IEC17025 : 2017

No. 27, F-2, 1st Floor, 2nd Street, Varalakshmi Nagar, Maduravoyal,  
(Opp. MGR Engineering College), Chennai - 600 095.

Ph.: 044 - 23780211, Cell : 93802 66480 / 86958 18108 / 90032 77250

E-mail: mkbestcalibration@gmail.com, www.mkbestcalibrationservices.com



CC-3340

## CERTIFICATE OF CALIBRATION

FF/7.8/01	Page No			1 of 1	
ULR No	CC334023000007859F	Date of Calibration	22.09.2023	Date of Receipt	22.09.2023
Certificate No	MKBL/23/09/0881-001	Due Date of Calibration	21.09.2024	Date of Issue	25.09.2023

CUSTOMER INFORMATION	DETAILS OF UNIT UNDER CALIBRATION	
M/S. KB DIAGNOSTICS., VANDAVASI.	Description	MICROPIPETTE 3
	Make / Model	P FACT PIPETTE / VA-500
	Range/Resolution	10 to 100 µl / 0.5 µl
	Serial No	396435
	Identification No	KBD/CB/GEN/04
	Location	----
	Calibrated at	LAB

STANDARD INSTRUMENTS DETAILS (The Standards Used are Traceable to National /International Standards)				
S.No	Description	Id.No/Sl. No	Certificate No	Validity
01	ELECTRONIC SEMI MICRO BALANCE	MK/CAL-96/477904	TVCSPL 23/07/638-01	20.07.2024
02	DIGITAL BAROMETER	MK/CAL-143	CC287922000005116F	12.10.2023
03	DIGITAL PEN TYPE THERMOMETER	MK/CAL-160	CRMTL/01/423101578-A4	03.07.2024

ENVIRONMENTAL & DUC CONDITIONS		REFERENCE STANDARD	
Temperature	23.5 °C	Reference Std	ISO 8655-6:2002
Humidity	53 % RH	Procedure No	MKBCS - MBV - 03
Condition of UUC Receipt	Good		

### CALIBRATION RESULTS


#### I.VOLUME CALIBRATION

S.No	UUC Reading (Mean) µl	STD Reading (Mean) µl	Deviation µl	Expanded Uncertainty (±) µl
1	20	20.00	0.00	1.18
2	40	40.00	0.00	1.18
3	60	60.05	-0.05	1.18
4	80	80.05	-0.05	1.18
5	100	100.05	-0.05	1.18

#### Remarks :

- The Expanded Uncertainty Associated with the Results is Calculated at a Confidence Level of Approximately 95% with a Coverage factor of K=2.
- The Calibration Certificate Shall not be Reproduced Except In Full, Without Written Approval Of The Laboratory.
- The Recalibration Interval Should be Determined on the User Requirement.
- The Results Stated In This Certificate Relate Only to the Item Calibrated.
- The User Should Determine The Suitability Of The Instrument For Is Intended Use.
- Resulted Volume Convert at 27°C of Water Temperature.
- Expanded Uncertainty is also Included Correction Factors.

Calibrated by

  
S.Chandrabose  
(Calibration Engineer)

x-x-x-x-Id Of Certificate-x-x-x-x



Authorised By

  
L.Magesh  
(MD/QM)