



Vanavil Calibrations

Make It Simple, But Significant!

CALIBRATION CERTIFICATE

Page no: 1 of 1

Customer Name & Address: M/S GOVERNMENT URBAN PRIMARY HEALTH CENTRE, METTUR - 636 456	Certificate No	VCPL/CC/2023/053-002
	Date of Calibration	21-10-2023
	Date of issue	25-10-2023
	Date of Receipt:	21-10-2023
	Next cal Date	20-10-2024

Details of Device Under Calibration

Description:	MICROPIPETTE	Make:	THERMO SCIENTIFIC
Range:	10 - 100 μ l	Model/Type:	FINNPIPETTE F3
Resolution:	0.2 μ l	Condition of DUC	GOOD
Serial Number	RW03154	Location:	LABORATORY
ID number	-	Cal. At:	LAB

Environmental Conditions

Temperature	23.5 $^{\circ}$ C	Calibration Method Used	ISO 8655 2002 PART-6
Humidity	51 % Rh	Cal procedure no:	-

Calibration Method Used

Master Details

SL no	Description	ID No / SL No	Certificate No	Make/Model	Valid Till
1	ELECTRONIC BALANCE	16031172	E-150323-1	WINTECH	15-03-2024

RESULT OF CALIBRATION

S NO	Duc Readings (μ l)	Standard Reading Mean Value (mg)	ERROR (mg)
1	10.0	10.02	-0.02
2	50.0	49.95	0.05
3	100.0	99.89	0.11


Remarks:

- 1 DUC is Defined as Device Under Calibration.
- 2 Calibration Frequency decided by the customer.
- 3 The above instrument is working in good condition.

Calibrated By:

R.GOPALAKRISHNAN

(CALIBRATION ENGINEER)



Authorised by:

M. KANNAN,
Technical Manager
Vanavil Calibrations Pvt Ltd.

*** End of the Report ***

Vanavil Calibrations Pvt. Ltd

(A Unit of Rainbow Bio Medical Engineers)

S2, Rishi Maharaj Apartment, Indra Nagar, 250 Reddiyur Main Road, Narasothipatty, Salem - 636 004.

No. 96,97,98, Anadhamaya Apartments, 2nd Floor, KVR Flats, 2nd Cross Street, Thangam Nagar, Pallikaranai, Chennai - 600 100.

Cell: 99949 13872 / 99949 23872 Ph: 0427-7965376

vanavilcalibrations@gmail.com www.rainbowbiomedicalengineers.com | www.vanavilcalibrations.com