S.M. ENTERPRISES

ISO 9001 : 2015 RQ91/3144

(Calibration and Testing of Electrical, Electronic & Precision Instruments)

27th Years of Service

Office/Cal. Lab : Plot No. 12, "Ganesh Niwas", Sanjeevaiah Nagar, Co-operative Society, Beside Tadbund Hanuman Temple, Sitafal Village, Secunderabad - 500 009. T.S. INDIA. Cell: 9848220798, 9948291008, 9948222008, 9948283008, 9948294008 Email: msutheer_98@yahoo.com, msutheer_98@smenterprises.org, SME_2008@YAHOO.IN Url: www.smenterprises.org

CALIBRATION CERTIFICATE

CALIBRATION ON 16/10/2023	CALIBRATION DUE 15/10/2024
1. Calibration Report Ref. No. & Date :	SME/SUJ/HYG/CAL-03/2023,Dt.16/10/2023
2. Requested by : (Name and address of the organization)	M/s. ICTC GHRIMS Srikakulam GHRIMS Srikakulam
3. Instrument Name	Digital Thermo Hygro Meter With probe
4.Make/Model:	HTC 2
5.Instrument ID No:	ICTC/HYG-03
6. Range :	0 to 50 ° C & 30 to 90 %
7. Accuracy :	+/-2° C & 5% RH
8. Environmental Conditions :	Temp/RH 25 +/- 2Deg C, RH - 65 +/- 5 %
9.Test parameter & condition :	Temperature & RH V/S Temperature& RH
10.Standard Instruments used for Calibration.	Digital Temp Humidity Meter
11. Traceability Certificate Ref. No.	C-230821-01-2,Dt.31/08/2023,Dued.31/08/2024

TEST RESULTS					
Sl No	Standard Reading Of Temperature & RH		Observed Reading Of Under Calibration Hygrometer in Deg C & %RH		Error Value +/-
	Temperature in °C	RH in %	Temperature in °C	RH in %	Temp in °C / RH in %
1	10.0	30	10.1	29	0.1 / 1
2	20.0	50	19.9	50	0.1 / 0
3	30.0	60	30.2	58	0.2 / 2
4	40.0	80	39.7	81	0.3 / 1
5	50.0	90	50.0	88	0.0 / 2

PLEASE NOTE: THE INSTRUMENTS ERROR FOUND TO BE WITHIN THE SPECIFIED ACCURACY LIMIT AND THIS CERTIFICATE VALID FOR PERIOD OF ONE YEAR FROM THE DATE OF ISSUE.

CALIBRATED BY
S.M.ENTERPRISES

ENGINEER INCHARGE

S.M. ENTERPRISES An ISO 9001-2015 Co. QA Incharge No. RQ91/3144 SECUNDERABAD

Traceability

CALIBRATION CERTIFICATE

CALIBRATION SERVICES

Signature valid