



Asanur
today at 4:06 PM



G MULTI SERVICES

78, Yehook Street, Gobichettipalayam, Erode Dist, Tamil Nadu. Pin: 638 452.
gmultiservicesgobi@gmail.com. Cell : 96597 20654.
Bio Medical Instruments Services

Er. A. Rathul Rasig DECE., BE (ICE)
Bio Medical Engineer

Calibration Protocol & Report for Medical Equipment

Equipment : PIPETTE
MAKE : FINNPIPEPTE F3
MODEL :100-1000µL

PROTECT CLASS /TYPE : Risk Class II
S. No. : AS/PI/002
Location : GPHC, ASANUR

S.No	Measurement Parameters	Status	Remarks
1	Electrical Test (a) Safety Standards (range : 210-240 V) (b) Equipotential ground & Earthing (c) Risk of shock/fire Hazards, if any (d) Chassis & Insulation Resistance (e) Earth & Chassis Leakage Current (f) Battery conditions & charging status	---	-
2	Physical Condition Test (a) Cabinet & visible damages, if any (b) Risk of mechanical damages, if any	No No	Checked OK
3	Performance Test (c) Mean 99.6µl (f) A -0.5 µl (g) Acc % -0.71 (h) CV% -0.16	OK OK OK Ok	
4	Calibration Test min volume 100 µl max Volume 1000 µl (e) Avar - 1.5, 5 µl (f) S 1.5 µl max Vol. 1000 µl (g) Accvar - 1.50, 0.50 (h) CV% 0.30, 0.10	Ok Ok Ok Ok	Tested
Formulated By: Edition Research Laboratory		Issued By:	User Department:
Endorsed By: Research team head			
Reviewed On: 22.09.23 Next Review : 21.09.24			

