



## CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/14204	MEDICAL DEVICES	Page No:1 of 1
Issue Date	12-10-2023	
SRF No & Date	SRF/23/00867-0005 & 11-10-2023	
Receipt Date	11-10-2023	
Calibration Date	11-10-2023	
Calibration Due	10-10-2024	

### Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE,  
MINJUR-601203, THIRUVALLUR DISTRICT.

### Details of Device Under Calibration (DUC)

Description : SEMI AUTO ANALYZER	Make & Model : MINDRAY & BA88A
Range : MULTI	Sr. No : WR09035500
Resolution : MULTI	Identification No : NA
DUC Condition : SATISFACTORY	Location : LABORATORY

### Environmental Conditions & Calibration Procedure Details

Environmental Details	Temperature: 25.6 ° C	Relative Humidity	52% RH
Calibration Procedure No	SBS/CP/MD/20	Calibration done at	ONSITE

### Reference Standards Details

S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

### ELECTRICAL SAFETY


#### RESULTS


S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	<b>Insulation Resistance</b> >20MΩ	Measured values in MΩ 97	Uncertainty in % ( ± ) 13.92
2	<b>Earth Leakage</b> <5000µA for B,BF,CF	Measured values in µA 125	Uncertainty in % ( ± ) 8.0
3	<b>Enclosure Leakage</b> <500µA for B,BF,CF	Measured values in µA 235	Uncertainty in % ( ± ) 7.2

#### REMARKS


1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
2. The user should determine the suitability of the instrument for its intended use.
3. The recalibration interval should be determined on the user requirement.
4. The results stated in this certificate relate only to the item calibrated.
5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

  
(Calibration Engineer)  
P. PRASANNA



Authorised Signatory

  
Technical Manager  
C. SHANMUGARAJ

Chief Executive