


06-10-2023 15:11:00

12.9250529

79.139629



MONTHLY SUMMARY

Name: Govt. UPHC, Subramaniyar Street, Saidapet
 PIN: 9295
 Month: January
 Year: 2023
 Student Group: Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

Click on the analyte to view Graphical Data

Date of Result Entered : 20/01/2023

Date of Report Published : 09/02/2023

Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	L
					CV	SD			
GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1890	119.87	6.38	7.65	103 mg/dL	-2.20	0.3
UREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1865	2.21	10.02	0.22	1.9 mg/dL	-1.40	0.6
CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1794	112.14	6.03	6.76	124 mg/dL	1.75	0.3

SDI Range	Interpretation
1 -1.00 to +1.00	Excellent.
1 ±1.01 to ±2.00	Good.
1 ±2.01 to ±2.99	Accept with caution. Warning Signal.
1 ±3.0	Unacceptable performance. Action Signal.

Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Summary	Change Password	Home	Sign-out

Site Maintained by

Swiftware Solutions

Best Viewed at pixels 1024

D. Small

MEDICAL OFFICER
UPHC SAIDAPET
VELLORE CITY MUNICIPAL CORPORATION