

2023 15:11:26

50377

96358



VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. UPHC, Subramaniyar Street, Saidapet  
 Lab No: 9295  
 Month: March  
 Year: 2023  
 Constituent Group: Chemistry III

[Details About Robust Analysis](#)  
[Detail About Monthly Summary](#)  
[Detail about SDI](#)

Click on the analyte to view Graphical Data

Date of Result Entered : 20/03/2023

Date of Report Published : 04/04/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1878	91.48	7.46	6.82	102 mg/dL	1.54	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1851	1.33	9.08	0.12	1.4 mg/dL	0.58	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1761	110.93	5.23	5.80	112 mg/dL	0.18	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

Site Maintained by  
**Swiftware Solutions**

Best Viewed at pixels 1024

*D. Anall*  
 MEDICAL OFFICER  
 UPHC SAIDAPET,  
 VELLORE CITY MUNICIPAL  
 CORPORATION