



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - APRIL 2023



PC-1024

Lab Name **Govt. Block PHC, Subbalapuram.M** Lab No **7216**
Constituent Group **Chemistry III** Date of Result Entered : **20/04/2023**
PT item **Lyophilized human serum based** Date of Report Published : **06/05/2023**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	115 mg/dL	-0.62	0.34
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	169	42.32	10.68	4.52	41 mg/dL	-0.29	0.70
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2 mg/dL	-0.91	0.01
4	URIC ACID III	Enzymatic / Uricase Colorimetric III	Robonik (Semi automatic / Fully Automatic)	38	4.58	9.59	0.44	5.1 mg/dL	1.18	0.14
5	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	105 mg/dL	-1.14	0.28
6	TRIGLYCERIDE III	GPO-PAP / Enzymatic Colorimetric / End Point III	Robonik (Semi automatic / Fully Automatic)	13	114.97	6.80	7.82	152 mg/dL	4.74	4.34

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. Block PHC, Subbalapuram.M
THE BLOCK MEDICAL OFFICER, THENI DISTRICT
Subbalapuram.M
TAMILNADU625517

Pamela Christudoss

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****