



CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/13071

Page. No : 1 of 1

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE, T.U.MANGAI-623533,KEELASEETHAI,RAMANATHAPURAM DISTRICT.	SRF No.	SRF/23/00664-0005
	SRF Date	05-10-2023
	Date of Receipt	04-10-2023
	Date of Calibration	05-10-2023
	Due Date for Calibration	04-10-2024
	Issue Date	06-10-2023

Details of Unit Under Calibration

Description	MICRO PIPETTE	Make	THERMO SCIENTIFIC
Range	100-1000µl	Model	FINNPIPETTE F3
Resolution	1µl	Material	PVC
Serial Number	PW16768	Operating Range	100-1000µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition				Calibration Method Used	
Temperature (°C)	23.5	Humidity (%RH)	52	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

SI. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSPL22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319


Result of Calibration in µl										
Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
1	100	99.94	99.96	99.94	99.92	99.93	99.95	-0.05	0.02	0.47
		99.94	99.95	99.96	99.97	99.98				
2	500	499.85	499.86	499.87	499.89	499.86	499.87	-0.13	0.02	0.47
		499.86	499.87	499.89	499.85	499.89				
3	1000	999.96	999.97	999.98	999.96	999.95	999.96	-0.04	0.01	0.47
		999.98	999.94	999.96	999.95	999.95				

Remarks


- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
5. Equipment used for Calibration were calibrated & traceable to National & International Standards
- 6.The indicated uncertainties are expanded uncertainty estimated for confidence level of approximately 95% for a coverage factor k=2.00 .
7. Calibration Liquid Used: Distilled or Deionized water conforming to Grades as specified in ISO 3696.

Calibrated By,


 (Calibration Engineer)
 MADHIBAN



Authorised by:


 (Quality Manager/Chief Executive)
 C.SIVABALAN

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
 E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.