

		CALI	<u>BRA</u>	<u>TION</u>	CERT	TIFICATI	<u> </u>		
CERTIFICATE NO:	SBS/CL/2:	3/13068			M	IECHANICAL	P	age No:1 of 1	
Issue Date	<u> </u>		_	06-10-2	023				
SRF No & Date				SRF/23/00664-0002 & 05-10-2023					
Receipt Date				05-10-2023					
Calibration Date				05-10-2023					
Calibration Due				04-10-2024					
Customer Name & Ad	dress_								
GOVERNMENT PRIMA	ARY HEALTH	CENTRE,			-				
T.U.MANGAI-623533,	KEELASEETHA	I,RAMANA	THAPUR	AM DIST	RICT.				
		De	tails of D	evice Un	der Calibra	ation (DUC)			
Description :	CENTRIFL	GE		Make &	Model	:	MICROSIL & MK-	-10	
Range :	4200	RPM		Sr. No		:	310/385		
Resolution :	100	RPM		Identific	ation No	:	NA		
DUC Condition :	Satisfactor	/		Location		:	LABORATORY		
	Envi	onmental (Condition	ns & Star	ndard Oper	rating Procedu	re Details		
Environmental Details Temperature:			e: 24.3 º	: 24.3 °C Relative Hu		umidity	54 % Rh		
Calibration Procedure No SBS/CP/ML/			/04	04 Calibration		done at ONSITE			
			Refer	ence Sta	ndards De	tails	•		
S.No Description	Description Make/ S			SI No:		Certificate No		Validity	
1 Digital Tachometer Lutron &			R DT-2238		JRPM-CCTR-/	JRPM-CCTR-A&S-2022-0036			

CALIBRATION RESULTS

<u>OALIDIA HON REGOLIO</u>							
S.No	DEVICE UNDER CALIBRATION READINGS	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY (±)			
-	RPM	RPM	RPM	%			
1	1000	999.9	0.1	5.8			
2	2000	1998	2	5.8			
3	3000	2999	1	5.8			

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately '95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

(Calibration Engineer)

K.SATHYAMOORTHY

Authorised Signatory

Chennai

Chennai

Chennai

Chennai

CCISIVABALAN)

Authorised Signatory

Chennai

CCISIVABALAN

SUNSHINE BIOMEDICAL SOLUTIONS