

## VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. UPHC, Kurichi	<a href="#">Details About Robust Analysis</a>
Lab No	9733	<a href="#">Detail About Monthly Summary</a>
Month	June	<a href="#">Detail about SDI</a>
Year	2023	
Constituent Group	Chemistry III	
<a href="#">Click on the analyte to view Graphical Data</a>		<a href="#">All Analyser Result</a> <a href="#">Print</a>
Date of Result Entered : 20/06/2023		
Date of Report Published : 04/07/2023		

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	<a href="#">GLUCOSE III</a>	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1823	286.63	5.17	14.82	378 mg/dL	6.17	0.69
2	<a href="#">CREATININE III</a>	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1848	2.57	6.64	0.17	2.9 mg/dL	1.93	0.01
3	<a href="#">CHOLESTROL III</a>	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1804	113.73	4.46	5.07	129 mg/dL	3.01	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>