

				CALIB	RAT	<u> 101</u>	N CERT	TIFIC	<u>A 7</u>	<u>'E</u>		
CERTIFICATE NO: SBS/CL/23/15933			3/15933	MECHANICAL			L	Page No:	1 of 1			
Issue Date			06-10-2023									
SRF N	o & Date				SRF/23/00626-0004 & 05-10-2023							
Receip	t Date				05-10-2023							
Calibra	ation Date				05-10-2023							
Calibration Due			04-10-2024									
Custor	ner Name &	Addres	38									
GOVE	RNMENT U	PGRA	DED PRIMA	RY HEALTH	CENTRE	,						
P.KOE	IKULAM,RA	TANMA	HAPURAM	DISTRICT.								
				Detai	ils of De	vice l	Jnder Calibr	ation (D	UC)			
Descri	ption	:	CENTRIFL	JGE		Make	e & Model	:		SANS LAB & NA		
Range	1	:		3500 RPM		Sr. N	lo	:		NA		
Resolu	ution	:		1 RPM		Ident	tification No	:		PKphc/centf/001		
DUC (Condition	_ :	Satisfactor	у		Loca	tion	:		LABORATORY	<u>-</u>	
			Envir	onmental Co	nditions	s & S1	tandard Ope	rating P	госес	dure Details		
Environmental Details Temperature: 25.5°C Relative Humidity 5.				52%Rh								
Calibration Procedure No SBS/CP/ML/			04 Calibration		one at ONSITE		ONSITE					
.,					Refere	nce \$	Standards D	etails		•		
S.No	No Description			Make/ SI No:			Certificate No				Validity	
1 Digital Tachometer			LINE SEIKI / 175-0034V			JRPM-CCTR-A&S-2023-0013				09-06-2024		

CALIBRATION RESULTS

S.No	DEVICE UNDER CALIBRATION	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY		
	RPM	RPM	RPM	%		
1	1000	999.0	1.0	4.2		
2	2000	1999.6	0.4	4.2		
3	3500	3499.6	0.4	4.2		

REMARKS

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.

Chennai

600 032

- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.

6. Equipment used for Calibration were calibrated & Maccalla to National & International Standards.

Calibrated By

(Calibration Engineer) N.VIJAY Authorised Signatory

(Quality Manager)
(C.SIVABALAN)

(Chief Executive)

SUNSHINE BIOMEDICAL SOLUTIONS