

				CAL	IBRAT	<u> 101</u>	CER	TIFICA	TE		0.500,4
CERTIFICATE NO: SBS/CL/23/15867					MEDICAL DEVICES					No:1 of 1	
Issue Date						27-10-2023					
SRF No & Date				SRF/23/00963-0001 & 26-10-2023							
Receipt Date					26-10-2023						
Calibration Date				26-10-2023							
Calib	ration Due	16				25-10-2024					
Custo	mer Name	& Addre	ss	2 20 0000000000000000000000000000000000					NST 18 300		14000
GOV	ERNMENT	PRIMA	RY HEA	LTH CENTR	Ē,	7 7 7 7 7 7		1333333333			
BERI	GAI-63510	5,KRISH	INAGIR	DISTRICT.							
					Details of De	evice L	Inder Calibra	ation (DUC)			
Descr	iption		SEMI A	UTO ANALY	ZER	Make	& Model	4	ROBONIK & PR	RIETEST TOUC	Э
Range	9	:	MULTI			Sr. N	0	:	ATCD0370519F	RBK	
Resol	ution	:	MULTI			Identi	fication No	:	NA		
DUC (Condition	:	SATIS	FACTORY		Locat	tion		BIO CHEMISTE	RY LAB	
			3 49	Environ	mental Cond	itions	& Calibration	n Procedure	Details		
Enviro	nmental De	tails		Temperature	:25.6°C		Relative Hui	midity	52% RH		
Calibr	alibration Procedure No SBS/CP/M		SBS/CP/MD/	/20		Calibration done at		ONSITE			
		- 72.87.9	38.17		Refere	ence S	tandards De	tails		* - 31	
S.No	Descriptio	n			Make/ SI No	;		Certificate	No		Validity
1 Electrical Safety Analyser		Rigel Medical & 44L-1059		M-230809-16-4			10-08-2024				

ELECTRICAL SAFETY

RESULTS

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)	
4	Insulation Resisitance	Measured values in MΩ	Uncertainty in % (±)	
	>20MΩ	87	13.92	
2	Earth Leakage	Measured values in µA	Uncertainty in % (±)	
	<5000µAfor B,BF,CF	195	9.3	
3	Enclosure Leakage	Measured values in µA	Uncertainty in % (±)	
	<500µAfor B,BF,CF	233	8.5	

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards. **Authorised Signatory**

Calibrated By

M.DINESH





Chief Executive