

			CALIBRA	ATION	CERTI	FICATE		
CERTIFICATE NO: SBS/CL/23/13148			3/13148	ME		CHANICAL	Pa	age No:1 of 1
Issue Date				07-10-2023				
SRF No & Date				SRF/23/00683-0001 & 06-10-2023				
Receipt Date				06-10-2023				
Calibration Date				06-10-2023				
Calibration Due				05-10-2024				
Custo	mer Name &	Address						
GOVE	RNMENT PR	MARY HEALTH	CENTRE,					
ARIYA	NKOTTAI,RA	MANATHAPURA	M DISTRICT.					
			Details of	f Device Un	der Calibrati	ion (DUC)		
Description : CENTRIFUGE		UGE	Make &	Model	:	B.D.INSTRUMENT	ATION & BDI-152	
Range :		: 3500	3500 RPM		Sr. No		NA	
Resolution : 1 RPM		RPM	Identification No		1	NA		
DUC C	DUC Condition : Satisfactory		ry	Location		:	LABORATORY	
		Env	ironmental Condit	tions & Star	ndard Opera	ting Procedu	re Details	
Environmental Details Temperature			Temperature: 25	e: 25.4°C Relative Hu		midity	54% Rh	
Calibration Procedure No SBS/CP/ML/			SBS/CP/ML/04	04 Calibration		done at ONSITE		
			Re	eference St	andards Deta	ails	•	
S.No	Description			Make/ SI No:		Certificate No		Validity
1	Digital Tachometer			LINE SEIKI / 175-0034V		JRPM-CCTR-A&S-2023-0013		09-06-2024

## CALIBRATION RESULTS

S.No	DEVICE UNDER CALIBRATION	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY	
	RPM	RPM	RPM	%	
1	1000	998.6	1.4	4.2	
2	2000	1998.6	1.4	4.2	
3	3000	2998.7	1.3	4.2	

## REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By
(Calibration Engineer)
E.ESWAR



Authorised Signatory					
Quality Manager (C.SIVABALAN)	Chief Executive				

## SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087, E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.