







CALIBRATION / TESTING REPORT					
CUSTOMER DETAILS : GOVERNMENT PRIMARY HEALTH CENTRE NAVALODIS, TIRUNELVELI.		CERTIFICATE NO		TICS/2023/NLS/TVL/28	
		EQUIPMENT NAME		PIPETTE	
		MANUFACTURE		MICROLUX	
		MODEL		100-1000 ML	
CALIBRATION DATE	07/10/2023	SERIAL NO		NA	
CALIBRATION DUE	06/10/2024	DEPARTMENT		LAB	
1. Have more dust in machine outside <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 2. Any Physical damaged in machine <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 3. Machine inside Cleaning (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No					
TEST EQUIPMENT	MANUFACTURE	CERTIFICATE NUMBER		VALIDITY	
ELECTRONIC SEMI MICRO BALANCE	SHIMADZU	TVCSP122/07/884-01		27-JUL-2023	
TEST REPORT					Temperature 25°C Humidity RH 55%
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	µL	100	100.13	± 1	PASS
		1000	1000.11	± 1	PASS
* The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipments having traceability to National/International standards.					
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE					
Calibrated By :		Authorized Sign With Seal			
		 			

No.763/19, Block No:10, KTC Nagar, Kulasekarapatti, Tirunelveli, Tamil Nadu, India - 627806.

☎ 04633-210998 ☎ +91 72000 04781, +91 95001 13759

■ transindiacalibration@gmail.com @ www.transindiacalibration.com

CIN: U52100TN2020PTC139842




CALIBRATION / TESTING REPORT						
CUSTOMER DETAILS :			CERTIFICATE NO	TICS/2023/NLS/TVL/27		
			EQUIPMENT NAME	MICROPIPETTE		
GOVERNMENT PRIMARY HEALTH CENTRE NAVALADI, TIRUNELVELL			MANUFACTURE	THERMOSCIENTIFIC		
			MODEL	10-100 µL		
CALIBRATION DATE	07/10/2023		SERIAL NO	QW13616		
CALIBRATION DUE DATE	06/10/2024		INSTALLED DEPARTMENT	LAB		
<p>1. Have more dust in machine outside <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p> <p>2. Any Physical damaged in machine <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p> <p>3. Machine inside Cleaning (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p>						
TEST EQUIPMENT	MANUFACTURE	CERTIFICATE NUMBER	VALIDITY			
ELECTRONIC SEMI MICRO BALANCE	SHIMADZU	TVCSP122/07/884-01	27-JUL-2023			
TEST REPORT					Temperature 25°C Humidity RH 55%	
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS	
VOLUME	µL	10	10.1	± 1	PASS	
		99	99.3	± 1	PASS	
* The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipments having traceability to National/International standards.						
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE			<input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED		<input type="checkbox"/> REMOVED FROM IUSE	
Calibrated By :			Authorized Sign With Seal			
						
						

No.763/19, Block No:10, KTC Nagar, Kulasekarapatti, Tirunelveli, Tamil Nadu, India - 627806.

© 04633-210998 ☎ +91 72000 04781, +91 95001 13759

■ transindiacalibration@gmail.com ● www.transindiacalibration.com

CIN: U52100TN2020PTC139842

CALIBRATION / TESTING REPORT					
CUSTOMER DETAILS : GOVERNMENT PRIMARY HEALTH CENTRE NAVALADI, TIRUNELVELI.			CERTIFICATE NO TICS/2023/NLS/TVL/29		
			EQUIPMENT NAME MICROPIPETTE		
CALIBRATION DATE 07/10/2023 CALIBRATION DUE DATE 06/10/2024			MANUFACTURE MICROLUX		
			MODEL 10-100 µL		
			SERIAL NO NA		
			INSTALLED DEPARTMENT LAB		
1. Have more dust in machine outside <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 2. Any Physical damaged in machine <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 3. Machine inside Cleaning (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No					
TEST EQUIPMENT		MANUFACTURE	CERTIFICATE NUMBER	VALIDITY	
ELECTRONIC SEMI MICRO BALANCE		SHIMADZU	TVCSPL22/07/884-01	27-JUL-2023	
TEST REPORT					Temperature 25°C Humidity RH 55%
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	µL	10	10.10	± 1	PASS
		100	99.78	± 1	PASS
*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipments having traceability to National/International standards.					
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE <input type="checkbox"/> REMOVED FROM USE					
Calibrated By : 				Authorized Sign With Seal 	

No.763/19, Block No:10, KTC Nagar, Kulasekarapatti, Tirunelveli, Tamil Nadu, India - 627806.

☎ 04633-210998 ☎ +91 72000 04781, +91 95001 13759

✉ transindiacalibration@gmail.com 🌐 www.transindiacalibration.com

CIN: U52100TN2020PTC139842