

# HOPE ULTRA SYSTEMS

No: 2/79, MAIN ROAD, MAMSAPURAM (PO),  
SIVAKASI, VIRUDHUNAGAR-626124.  
E-mail : [hopeultrasystems@gmail.com](mailto:hopeultrasystems@gmail.com)

GSTIN. 33DLMPK3057D1Z9  
Mobile: +919487402761, +919715223192.

## CALIBRATION CERTIFICATE

Requested By  
**GPHC, THIRUVENGADAM.**

| <b>Scope:</b>  |                                     | Asset no            | : PHC/TVM-LAB/002 |
|----------------|-------------------------------------|---------------------|-------------------|
| Equipment Name | : Micropipette                      | Calibration Done at | : customer place  |
| Make           | : Thermo scientific(Finnpipette F3) | Calibration Date    | : 10/09/2023      |
| Model          | : 100-1000 $\mu$ l                  | Next call due on    | : 09/09/2024      |
| Sl no          | : RW11907                           |                     |                   |

### TEST RESULTS Volume $\mu$ l

| Sl. No. | DUC reading( $\mu$ l) | Standard Reading ( $\mu$ l) |         | Correction ( $\mu$ l) |
|---------|-----------------------|-----------------------------|---------|-----------------------|
|         |                       | Trail1                      | Trail 2 |                       |
| 01      | 100                   | 100.0                       | 100.0   | +0                    |
| 02      | 200                   | 200.0                       | 200.0   | +0                    |
| 03      | 500                   | 500.0                       | 500.0   | +0                    |
| 04      | 800                   | 800.0                       | 800.0   | +0                    |
| 05      | 1000                  | 1000.0                      | 1000.0  | +0                    |

Error Acceptance : Not Applicable (user defined)

Equipment Status : Working in Good condition

Materials used for testing and calibration:

- Electronic weighing scale(EWC/SF400C)

Note:

- Calibration points were selected as per customer request.
- Each reading was taken by repeatability.
- Results reported are valid at the time of and under the stated conditions of measurements.

Hope Ultra Systems

Authorized signatory

