

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. UPHC, East Thambram
 Lab No 12259
 Month June
 Year 2023
 Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 19/06/2023

Date of Report Published : 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	260 mg/dL	-1.80	0.89
2	UREA III	Others (Urease Berthelot / Ned Dye) III	Any Analyser (Automation / Semi Automation)	50	33.99	14.82	5.04	43 mg/dL	1.79	1.42
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation)	67	2.58	6.39	0.16	2.5 mg/dL	-0.48	0.04
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	112 mg/dL	-0.34	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

E. Selvan
MEDICAL OFFICER
 Govt. Urban Primary Health Center
 Poondibazaar, East Tambaram,
 Chennai - 600 059.

RECEIVED JADIGEM
 19/06/2023 10:45:00 AM
https://home.cgovtore.ac.in/diagc/ViewMonthlySummary/viewResult2_new.aspx?resDetId=266&noSamples=1&month=6&year=2023&groupId=9&yearGroup=110