



External Quality A...  
home.cmcvellore.ac.in



VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. P.H.C. Kariyamannickam  
Lab No: 12757  
Month: August  
Year: 2023  
Constituent Group: Chemistry III

[Detail About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail About SDI](#)

Click on the analysis to view Graphical Data

[All Analysis Result](#)

[Print](#)

Date of Result Created : 19/08/2023


Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyser Name	No. of Participants	CV	Performance		User Value	SDI	Q
						CV	TD			
1	GLUCOSE (F)	SCO-PGD (F)	Robonik (Semi automatic - Fully Automatic)	1841	175.58	3.13	3.00	175 mg/dL	0.03	0.42
2	CREATININE (F)	JAFFES Kinetic - Alkaline Phosph (F)	Robonik (Semi automatic - Fully Automatic)	1852	1.76	6.77	6.12	2.1 mg/dL	2.85	0.91
3	CHOLESTROL (F)	CHOD-PAP (F)	Robonik (Semi automatic - Fully Automatic)	1764	113.80	4.07	4.04	114 mg/dL	0.04	0.22

SDI Range	Interpretation
Within -1.00 to -1.00	Excellent.
Within -1.01 to -2.00	Good.
Within -2.01 to -2.99	Accept with caution. Warning Signal.
Beyond -3.0	Unacceptable performance. Action Signal.

Links			
<a href="#">Monthly Summary</a>	<a href="#">Quality methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Weekly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign out</a>

Site Maintained by



Best viewed at pixels 1024 x 768

*S. Kishor*  
MEDICAL OFFICER  
Govt Primary Health Centre  
94 Kariyamannickam  
Mannachannallur (T) x1  
Truchirappalli (Dt) - 625 025

**VIEW LAB MONTHLY SUMMARY**

Lab Name      Govt. PHC, Kariyamattickam  
 Lab No        12725  
 Month         July  
 Year           2023  
 Constituent Group      Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 19/07/2023

Date of Report Published : 07/08/2023

Sl No.	Analyte	Reference Interval	Assay Method	No. of Participants	CV	Performance		Total Error	SDI	Q
						CV	SE			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	6.07	116 mg/dL	1.10	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1850	5.83	6.00	0.35	6.1 mg/dL	0.77	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.66	112 mg/dL	-0.21	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

**Links**

<a href="#">Home</a>	<a href="#">About Us</a>	<a href="#">Contact Us</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Privacy Policy</a>	<a href="#">Terms &amp; Conditions</a>	<a href="#">FAQ</a>	<a href="#">Helpdesk</a>

*J. Senthil*  
**MEDICAL OFFICER**  
 Govt Primary Health Centre  
 94, Kariyamattickam  
 Mannachanattur (Tn)  
 Tiruchirappalli (Dt) - 621 005



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - APRIL 2023



Lab Name: Govt. PHC, Kariyamanickam  
 Constituent Group: Chemistry III  
 PT item: Lyophilized human serum based

Lab No: 12725  
 Date of Result Entered: 17/04/2023  
 Date of Report Published: 06/05/2023

Sl No	Analyte	Method / Principle Name	Analyser Name	Total Participants	CV	Reference		Your Value	Diff	%
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1832	119.45	6.94	7.21	113 mg/dL	-8.89	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1820	2.16	6.09	0.16	2.3 mg/dL	0.89	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1750	111.61	5.19	5.79	194 mg/dL	-1.31	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS:  
 Govt. PHC, Kariyamanickam  
 KARIYAMANICKAM  
 Kariyamanickam  
 TAMILNADU621005

Coordinator Contact Details:  
 Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
 Contact Number: 0416-2283102

*Pamela Christudoss*  
 Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

*J. Srinivasan*  
 MEDICAL OFFICER  
 Govt Primary Health Centre  
 94 Kariyamanickam  
 Mannachannallur (Tal)  
 Truchirappalli (Dt) - 621 005

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential



External Quality A...  
home.cmcvellore.ac.in



VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Kariyambran

[Details About Request Analyser](#)

Lab No: 12725

[Detail About Monthly Summary](#)

Month: April

[Detail about SDI](#)

Year: 2023

Constituent Group: Chemistry II

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered: 11/04/2023

Date of Report Published: 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyser Name	No of Participants	CV	Participants		Your Value	SDI	Z
						TC	SC			
1	GLUCOSE II	GOE-POD II	Robtrak (Semi-automatic / Fully Automatic)	1825	2.63	5.17	14.82	279 mg/dL	-0.51	0.65
2	CREATININE II	Jaffee-Kolmer - Alkaline Phosphate II	Robtrak (Semi-automatic / Fully Automatic)	1846	2.57	6.64	0.57	2.7 mg/dL	0.76	0.01
3	CHOLESTROL II	CHOD-PAP II	Robtrak (Semi-automatic / Fully Automatic)	1804	113.73	4.46	3.67	1.3 mg/dL	-0.14	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent
Within +1.01 to +2.00	Good
Within +2.01 to +2.99	Accept with caution, Warning Signal
Beyond +3.0	Unacceptable performance, Action Signal

Links			
<a href="#">Weekly Address</a>	<a href="#">Weekly Address</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Weekly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Logout</a>



Site Maintained by  
**Swiftware Solutions**

Best Viewed at pixels 1024 X 768

*J. Senthil*  
MEDICAL OFFICER  
Govt Primary Health Centre  
04, Kariyambran (Kannur)  
Mannacilandi (M.T)  
Tiruchirappalli (01) - 620 005



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - MAY 2023



Lab Name Govt. PHC, Kariyamanickam  
Constituent Group Chemistry III  
PT Item Lyophilized human serum based

Lab No 12725  
Date of Result Entered : 18/05/2023  
Date of Report Published : 05/06/2023

SLNo	Analyte	Method / Principle Name	Analyser Name	No of Participants	CV	Participants		Total Value	SDI	Q
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	181 mg/dL	0.01	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	1.9 mg/dL	1.06	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83	118 mg/dL	0.58	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :  
Govt. PHC, Kariyamanickam  
KARIYAMANICKAM  
Kariyamanickam  
TAMILNADU621005

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

*[Signature]*  
MEDICAL OFFICER  
Govt Primary Health Centre  
94 Kariyamanickam  
Mannachannallur (Tk)  
Tiruchirappalli (Dt) - 621 005

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential