

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. UPHC, Anamanthapur
Lab No 10822
Month July
Year 2023
Constituent Group Chemistry II

[Details About Robust Analysis](#)
[Detail About Monthly Summary](#)
[Detail about SDI](#)

Click on the analyte to view Graphical Data

All Analyser Result

Date of Result Entered : 2007:

Date of Report Published : 07:

Sl No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants		Your Value
						CV	SD	
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.87	104 mg/dL
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	5.4 mg/dL
3	CHOLESTROL II	CHOD-PAP II	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	112 mg/dL

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Results
Yearly Summary	Change Password	Home	Sign-out

e. Suresh