

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. UPHC, Anumanthaputheri	Details About Robust Analysis
Lab No	16822	Detail About Monthly Summary
Month	August	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	

Click on the analyte to view Graphical Data

All Analyser Result P:

Date of Result Entered : 18/08/23

Date of Report Published : 01/09/23


Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value
						CV	SD	
1	GLUCOSE III	GOD-POD III	Rebonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	180 mg/dL
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Rebonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.3 mg/dL
3	CHOLESTROL III	CHOD-PAP III	Rebonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.84	117 mg/dL

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Mon
Yearly Summary	Change Password	Home	Sign-out

R. Dhanasekaran



Samsung Triple Camera
Sudha chenni D