

CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/16416		MECHANICAL		Page No 1 of 1
Issue Date		19-10-2023		
SRF No & Date		SRF/03/009/14-0004 & 17-10-2023		
Receipt Date		17-10-2023		
Calibration Date		17-10-2023		
Calibration Due		19-10-2024		
Customer Name & Address				
GOVERNMENT URBAN PRIMARY HEALTH CENTRE, ANUMANTHAPUTHERI-600002,CHENGALPATTU DISTRICT				
Details of Device Under Calibration (DUC)				
Description	CENTRIFUGE	Make & Model	REM & R-60	
Range	5250 RPM	Sr. No	ZEAN-01960	
Resolution	10 RPM	Identification No	NA	
DUC Condition	Satisfactory	Location	LABORATORY	
Environmental Conditions & Standard Operating Procedure Details				
Environmental Details		Temperature: 25.5°C	Relative Humidity	52%/rh
Calibration Procedure No		SBS/CP/ML/04	Calibration done at	
			ON-SITE	
Reference Standards Details				
S.No	Description	Make/Sl No:	Certificate No	Validity
1	Digital Tachometer	LINE SERVO / 175-0034V	JRPM-CCTR-AAS-2023-0613	06-06-2024

CALIBRATION RESULTS				
S.No	DEVICE UNDER CALIBRATION	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY
	RPM	RPM	RPM	%
1	1000	999.1	0.9	4.2
2	2000	1999.5	0.5	4.2
3	3500	3499.5	0.5	4.2

- REMARKS**
- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
 - The user should determine the suitability of the instrument for its intended use.
 - The recalibration interval should be determined on the user requirement.
 - The results stated in this certificate relate only to the item calibrated.
 - The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor $k=2.00$.
 - Equipment used for Calibration were calibrated & traceable to National & International Standards.


Calibrated By
P. Prasanha
(Calibration Engineer)
P.PRASANHA



Authorised Signatory
 (Quality Manager) (Chief Executive)
C. Sivabalan
(C.SIVABALAN)

SUNSHINE BIOMEDICAL SOLUTIONS
No.68, First Floor, Poromagal Main Road, Ekkattuthangal, Chennai-600 032. Tel. 044-2225 2087.
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.

P. Prasanha



Samsung Triple Camera
Sudha chenni D

Item No. 102