



**CHRISTIAN MEDICAL COLLEGE**  
DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - AUGUST 2023



Lab Name: Govt. UPHC, Meenakshi Maternity  
 Constituent Group: Chemistry III  
 PT Item: Lyophilized human serum based  
 Lab No: 10740  
 Date of Result Entered: 20/08/2023  
 Date of Report Published: 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	175 mg/dL	-0.06	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.8 mg/dL	0.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	110 mg/dL	-0.82	0.22
<b>SDI Range</b>										
				<b>Interpretation</b>						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

**LAB ADDRESS :**  
 Govt. UPHC, Meenakshi Maternity  
 DR. ANANDHAVIDHYA, CHIEF MEDICAL OFFICER, BHARATHIPARK ROAD - II, RAMALINGAM COLONY  
 Coimbatore  
 TAMILNADU641043

**Coordinator Contact Details:**  
 Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
 Contact Number: 0416-2283102

*Pamela Christudoss*  
 Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential  
 CMC EQAS does not sub contract any components  
 \*\*\*\*\* End of Report \*\*\*\*\*

*A. Aravindhan*  
**CHIEF MEDICAL OFFICER**  
 U-PHC, M.M. HOME  
 158, BHARATHI PARK ROAD-2  
 RAMALINGAM COLONY  
 COIMBATORE - 43