

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. UPHC, Meenakshi Maternity	Details About Robust Analysis
Lab No	10740	Detail About Monthly Summary
Month	July	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	

Click on the analyte to view Graphical Data

Date of Result Entered : 20/07/2023

Date of Report Published : 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	107 mg/dL	-0.39	0.28
2	UREA III	Others (Urease Berthelot / Ned Dye) III	Any Analyser (Automation / Semi Automation)	47	52.80	7.28	3.85	68 mg/dL	3.95	1.12
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	5.9 mg/dL	0.20	0.02
4	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	113 mg/dL	0.00	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

- [Modify Address](#)
- [Modify Methods](#)
- [Yearly Summary](#)
- [Submit Lab Results](#)
- [View Lab Monthly Summary](#)
- [Home](#)
- [Change Password](#)
- [Sign-out](#)

CHIEF MEDICAL OFFICER
 U-PHC, M.M. HOME
 158, BHARATHI PARK ROAD-2
 RAMALINGAM CO. QNY
 COIMBATORE 43