



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JUNE 2023



PC-1024

10740

20/06/2023

04/07/2023

Lab Name Govt. UPHC, Meenakshi Maternity

Lab No

Constituent Group Chemistry III

Date of Result Entered :

Test item Lyophilized human serum based

Date of Report Published :

No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	370 mg/dL	5.63	0.69
	UREA III	Others (Urease Berthelot / Ned Dye) III	Any Analyser (Automation / Semi Automation)	50	33.99	14.82	5.04	38.2 mg/dL	0.84	1.42
	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.7 mg/dL	0.76	0.01
	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	109 mg/dL	-0.93	0.24

SDI Range	Interpretation
in -1.00 to +1.00	Excellent.
in ±1.01 to ±2.00	Good.
in ±2.01 to ±2.99	Accept with caution. Warning Signal.
and ±3.0	Unacceptable performance. Action Signal.

ADDRESS :
UPHC, Meenakshi Maternity
MANDHAVIDHYA, CHIEF MEDICAL OFFICER, BHARATHIPARK ROAD - II, RAMALINGAM COLONY
Coimbatore
KARNATAKA
KARNATAKA 561043

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

A. Anandharya
CHIEF MEDICAL OFFICER
U-PHC, M.M. HOME
158, BHARATHI PARK ROAD-2
RAMALINGAM COLONY
COIMBATORE - 43