

10<sup>th</sup> Sept 2023

## Calibration Certificate

Name & Address of the customer	SUMAN HOSPITAL, SULTANPUR
Make	Transasia
Product Name	Erba EM 200 - Fully Automated (Biochemistry Analyzer)
Serial No./Instrument ID	B200712

This is to certify that the cell counter installed at the above mentioned customer site has been calibrated with (Lot # 41000092, Expiry Date: 22/09/2023)

Sl. No.	Procedures	Remarks
1	Startup	PASS
2	Reproducibility	PASS
3	QC Bio-Rad	PASS
4	Carryover	PASS
5	Calibration; Calibrator(Lot#410092)	PASS
6	Calibration Verification (c.f.a.s..Lot: 4714920001)	PASS

<b>Date of Calibration : 10/09/2023</b>	<b>Due date for Calibration: 09/09/2023</b>
-----------------------------------------	---------------------------------------------

For Transasia Bio-Medicals Ltd.,

  
Mr. Gobinda Acharya

(Sr. Manager Service Operation)

