












CALIBRATION / TESTING REPORT				
CUSTOMER DETAILS : GOVERNMENT PRIMARY HEALTH CENTRE SOORANAM, SIVAGANGAL.		CERTIFICATE NO	TICS/2023/SRM/SG/01	
		EQUIPMENT NAME	SEMI ANALYZER	
		MANUFACTURE	ROBONIK	
		MODEL	PRIETEST TOUCH	
CALIBRATION DATE	20/10/2023	SERIAL NO	AT2820211RBK	
CALIBRATION DUE DATE	19/10/2024	INSTALLED	LAB	
		DEPARTMENT		
1. General maintenance required <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 2. Any Physical damage observed <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 3. Spare Replacement (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No				
TEST EQUIPMENT	MANUFACTURE	MODEL	SERIAL NUMBER	
ELECTRICAL SAFTY ANALYZER	DATREND	V Pad-mini	VPM 18040011	
PARAMETER	UNITS	VALUE	LIMITS	STATUS
Main Voltage L1-L2	V	227.8	240±10	PASS
Main Voltage L1-G(PE)	V	226.7	240±10	PASS
Main Voltage L2-G(PE)	V	0.78	<5	PASS
Load Current	A	0.388	As per unit	PASS
Protective Earth	Ω	0	<0.3	PASS
Point-to-Point Resistance	Ω	0	<0.5	PASS
Equipment Leakage	μA	0.17	Class I - < 500 Class II - < 100	PASS
Applied Part Leakage	μA	0	Type BF - < 5mA Type CF- <50μA*	PASS
Note : Test Standard: IEC 62353 Defibrillator paddles 100 μA, per IEC 60601-2-4				
*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.				
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE				
Calibrated By :			Authorized Sign With Seal	
				
				

CALIBRATION / TESTING REPORT					
CUSTOMER DETAILS : GOVERNMENT PRIMARY HEALTH CENTRE SOORANAM, SIVAGANGAI.			CERTIFICATE NO		TICS/2023/SRM/SG/02
			EQUIPMENT NAME		CENTRIFUGE
			MANUFACTURE		BD INSTRUMENTATION
			MODEL		BDI153
CALIBRATION DATE		20/10/2023	ASSET NO		SRMPH002
CALIBRATION DUE		19/10/2024	DEPARTMENT		LAB
<p>1. General maintenance required <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p> <p>2. Any Physical damage observed <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p> <p>3. Spare Replacement (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p>					
TEST EQUIPMENT		MANUFACTURE		MODEL	SERIAL NUMBER
DIGITAL TACHOMETER		KUSAM MECO		KM2234 BL	S1049875
TEST REPORT					Temperature 25°C Humidity RH 55%
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
ROTATION PER MINUTE	RPM	SPEED 2	1883	± 5	PASS
		SPEED 5	2867		
*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipment having traceability to National/International standards.					
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE					
Calibrated By :			Authorized Sign With Seal		
					
					

CALIBRATION / TESTING REPORT					
CUSTOMER DETAILS : GOVERNMENT PRIMARY HEALTH CENTRE SOORANAM, SIVAGANGAL.		CERTIFICATE NO		TICS/2023/SRM/SG/03	
		EQUIPMENT NAME		MICROPIPETTE	
		MANUFACTURE MODEL		THERMO SCIENTIFIC FINE PIPETTE 100-1000 µL	
		ASSET NO		SRM PH 03	
CALIBRATION DATE	20/10/2023	INSTALLED DEPARTMENT		LAB	
CALIBRATION DUE DATE	19/10/2024				
1. Have more dust in machine outside <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 2. Any Physical damaged in machine <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 3. Machine inside Cleaning (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No					
TEST EQUIPMENT	ID. NO	CERTIFICATE NUMBER	VALIDITY		
ELECTRONIC SEMI MICRO BALANCE	NCS/WB01	TSC/23-24/5424-1	27-JUNE-2024		
TEST REPORT					Temperature 25°C Humidity RH 55%
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	µL	100	100.06	± 1	PASS
		1000	1000.02	± 1	PASS
*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipments having traceability to National/International standards.					
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE					
Calibrated By :		Authorized Sign With Seal			
					
					

CALIBRATION / TESTING REPORT					
CUSTOMER DETAILS :		CERTIFICATE NO		TICS/2023/SRM/SG/04	
		EQUIPMENT NAME		MICROPIPETTE	
GOVERNMENT PRIMARY HEALTH CENTRE SOORANAM, SIVAGANGAL.		MANUFACTURE		THERMO SCIENTIFIC	
		MODEL		FINE PIPETTE 10-100 µL	
CALIBRATION DATE	20/10/2023	ASSET NO		TM PH 04	
CALIBRATION DUE DATE	19/10/2024	INSTALLED DEPARTMENT		LAB	
<p>1. Have more dust in machine outside <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p> <p>2. Any Physical damaged in machine <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p> <p>3. Machine inside Cleaning (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No</p>					
TEST EQUIPMENT	ID. NO	CERTIFICATE NUMBER		VALIDITY	
ELECTRONIC SEMI MICRO BALANCE	NCS/WB01	TSC/23-24/5424-1		27-JUNE-2024	
TEST REPORT					Temperature 25°C Humidity RH 55%
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	µL	10	99.92	± 1	PASS
		99	99.98	± 1	PASS
*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipments having traceability to National/International standards.					
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE					
Calibrated By :			Authorized Sign With Seal		
					

CALIBRATION / TESTING REPORT					
CUSTOMER DETAILS : GOVERNMENT PRIMARY HEALTH CENTRE SOORANAM, SIVAGANGAI.		CERTIFICATE NO		TICS/2023/SRM/SG/05	
		EQUIPMENT NAME		PIPETTE	
		MANUFACTURE		ERBA BIOHIT	
		MODEL		20 µL FIXED	
CALIBRATION DATE	20/10/2023	ASSET NO		SRM PH 05	
CALIBRATION DUE	19/10/2024	DEPARTMENT		LAB	
1. Have more dust in machine outside <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 2. Any Physical damaged in machine <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No 3. Machine inside Cleaning (if required) <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No					
TEST EQUIPMENT		ID. NO		CERTIFICATE NUMBER	
ELECTRONIC SEMI MICRO BALANCE		NCS/WB01		TSC/23-24/5424-1	
TEST REPORT					Temperature 25°C Humidity RH 55%
PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	µL	20	20.02	± 1	PASS
*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipments having traceability to National/International standards.					
<input checked="" type="checkbox"/> ACCEPTABLE FOR USE <input type="checkbox"/> CORRECTIVE MAINTENANCE REQUIRED <input type="checkbox"/> REMOVED FROM USE					
Calibrated By : Authorized Sign With Seal					
