

# CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/13644	MEDICAL DEVICES	Page No: 1 of 1
Issue Date	07-10-2023	
SRF No & Date	SRF/23/00788-0001 & 07-10-2023	
Receipt Date	07-10-2023	
Calibration Date	07-10-2023	
Calibration Due	06-10-2024	

**Customer Name & Address**

GOVERNMENT PRIMARY HEALTH CENTRE,  
VADUGACHERY-611109,NAGAPATTINAM DISTRICT.

**Details of Device Under Calibration (DUC)**

Description	SEMI AUTO ANALYZER	Make & Model	ROBONIK & PRIETEST TOUCH
Range	MULTI	Sr. No	ATCD0200919RBK
Resolution	MULTI	Identification No	NA
DUC Condition	SATISFACTORY	Location	LABORATORY

**Environmental Conditions & Calibration Procedure Details**

Environmental Details	Temperature 25.4° C	Relative Humidity	52% RH
Calibration Procedure No	SBS/CP/MD/20	Calibration done at	ONSITE

**Reference Standards Details**

S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

**ELECTRICAL SAFETY**

**RESULTS**


S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resisitance >20MΩ	Measured values in MΩ 100	Uncertainty in % (±) 13.92
2	Earth Leakage <5000µAfor B,BF,CF	Measured values in µA 154	Uncertainty in % (±) 7.8
3	Enclosure Leakage <500µAfor B,BF,CF	Measured values in µA 224	Uncertainty in % (±) 7.3


**REMARKS**

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
2. The user should determine the suitability of the instrument for its intended use.
3. The recalibration interval should be determined on the user requirement.
4. The results stated in this certificate relate only to the item calibrated.
5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

Authorised Signatory

  
(Calibration Engineer)  
P.PRASANNA



Technical Manager  
C.SHANMUGARAJ

Chief Executive

## SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,  
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.